

October 1, 2008

Lief Organics  
29013 Avenue Penn  
Valencia, CA 91355

Attn: Adel Villalobos

Job No: 109041

LO

---

## LABORATORY REPORT

---

Samples Received: Six (6) Samples  
Date Received: 09/25/2008  
Purchase Order No: COD

The samples were analyzed as follows:

Analysis


Page

Lead for CA Prop 65 by SOP 7130, Rev 5

2 - 4

---

  
D.J. Northington, Ph.D.  
Technical Director

  
Samina N. Hussain  
Senior Chemist

This report is to be reproduced in its entirety.

Page 1 of 4

Lead for CA Prop 65 by SOP 7130, Rev 5  
Quantitative Analysis Report  
Isotope Dilution Inductively Coupled Plasma-Mass Spectrometry

<u>Sample ID</u>	<u>Lead</u>	Parts Per Million ( $\mu\text{g/g}$ )	<u>Specification</u>	<u>Pass/Fail</u>
Lot C08060 Mg Citrate Granular	ND		NMT 10	Pass
Lot D08036 Ascorbic acid	ND		NMT 2	Pass
Lot I08035 CaCitrate Gran	ND		NMT 5	Pass

Detection Limit: 0.02

Date Analyzed: 09-29-08

## Quality Control Summary

Sample: Lot I08035 CaCitrate Gran

<u>Analyte</u>	<u>Sample Result</u>	<u>Duplicate Result</u>	<u>Average Result</u>	<u>Sample RPD</u>	<u>Spike Conc</u>	<u>Spike Result</u>	<u>Spike % Rec</u>
Lead	ND	0.0244	NA	NA	5.00	5.03	101
Control Limits:				<25			80 - 120

Date Analyzed: 09-29-08

Lead for CA Prop 65 by SOP 7130, Rev 5  
Quantitative Analysis Report  
Isotope Dilution Inductively Coupled Plasma-Mass Spectrometry

<u>Sample ID</u>	<u>Lead (µg/g)</u>	<u>Avg. Capsule Wt. (g)</u>	<u>Lead (µg/Capsule)</u>
After Baby Boost Capsules #1 Lot 20926	0.084	0.9707	0.082
Detection Limit:	0.01		

Date Analyzed: 09-29-08

## Quality Control Summary

Sample: After Baby Boost Capsules #1 Lot 20926

<u>Analyte</u>	<u>Sample Result</u>	<u>Duplicate Result</u>	<u>Average Result</u>	<u>Sample RPD</u>	<u>Spike Conc</u>	<u>Spike Result</u>	<u>Spike % Rec</u>
Lead	0.0842	0.0845	0.0844	NA	2.56	2.61	99
Control Limits:				<25			80 - 120

Date Analyzed: 09-29-08

Lead for CA Prop 65 by SOP 7130, Rev 5  
Quantitative Analysis Report  
Isotope Dilution Inductively Coupled Plasma-Mass SpectrometryParts Per Million ( $\mu\text{g/g}$ )

<u>Sample ID</u>	<u>Lead</u>	<u>Specification</u>	<u>Pass/Fail</u>
Lot E08105 Sipernat 225 Precipitate Silica	0.260	NA	NA
Lot I08005 Mg Stearate	ND	NMT 10	Pass
Detection Limit:	0.02		

Date Analyzed: 09-29-08

## Quality Control Summary

Sample: Lot I08005 Mg Stearate

<u>Analyte</u>	<u>Sample Result</u>	<u>Duplicate Result</u>	<u>Average Result</u>	<u>Sample RPD</u>	<u>Spike Conc</u>	<u>Spike Result</u>	<u>Spike % Rec</u>
Lead	ND	ND	ND	NA	9.80	8.80	90
Control Limits:				<25			80 - 120

Date Analyzed: 09-29-08

# ANALYTICAL REQUEST FORM



Bodycote Testing Group • [bodycotetesting.com](http://bodycotetesting.com)  
 9240 Santa Fe Springs Road, Santa Fe Springs, CA 90670

562.948.2225  
 Fax 562.948.5850

Send Report To Contact: <u>Adel Villalobos</u> Company: <u>Lief Organics</u> Address: <u>29013 Avenue Penn</u> City: <u>Valencia</u> State: <u>CA</u> Zip: <u>91355</u> Phone: <u>661-775-2500</u> Fax: <u>661-775-2549</u>	Send Invoice To AP Contact: <u>Same</u> Address: Phone: Project: Purchase Order / CC:
---	--

PRECAUTIONS:	COMMENTS:
--------------	-----------

Turn-around time (business days):   
 Normal 7-10 days   
 Rush 5 days   
 Rush 3 days   
 Rush 1 day

#	Sample ID:	Matrix / Product:	Analysis(es):
1*	Mg Citrate Granular		Lead Prop 65
2*	Ca Citrate Gran		Lead Prop 65
3*	Mg Stearate		Lead Prop 65
4*	Supernat 225		Lead Prop 65
5*	Ascorbic Acid		Lead Prop 65
	After Baby Boost Capsules #1 Lot 20926		Lead Prop 65
6			
7			

09.25.08 SD: Also on sample:  
 1\* Lot: C08060  
 2\* Lot: I08035  
 3\* Lot: I08005  
 4\* Lot: E08105  
 5\* Lot: D08036

Relinquished by:	Company: <u>UPS</u>	Date: <u>9/24/08</u>	Time: <u>11:44</u>	Received by: <u>Sarah Davis</u>	Company: <u>Bodycote</u>
------------------	---------------------	----------------------	--------------------	---------------------------------	--------------------------

NOTES: Samples will be discarded 30 days after invoicing unless notified in writing or high hazard.  
 All services provided will adhere to the BODYCOTE TESTING GROUP Terms & Conditions.

Job No.: 109041